THE PERFORMANCE OF EUROPEAN CALIBRATION SERVICES REGARDING RADON IN AIR

- EMPIR Project *Metrology for Radon* –

Preliminary Results

T. R. Beck

German Federal Office for Radiation Protection (BfS)





EMPIR Project *Metrology for Radon*

Work Package 5.2 (extract): Radon Intercomparison

- Selection of a suitable reference instrument for the use as transfer standard and its preparation for the intercomparison (good linearity and repeatability of the measurements, a high measurement range, mechanical robustness, ease of use).
- Developing of a protocol for the comparison, including a form for the participants to document their calibration procedures and measures for quality assurance.
- Excecution of the comparison.



EMPIR Project *Metrology for Radon*

Work Package 5.2 (extract): Radon Intercomparison

 Assessment of the results of the intercomparison regarding their closeness of agreement (precision). Conclusions shall be drawn for the realization of radon activity concentration in air at the European radon calibration facilities in the range from 300 Bq/m³ to 10 000 Bq/m³.



Transfer comparison device: AlphaGUARD PQ 2000 PRO TTL





Transfer Comparison Device

Procedure for checking linearity and precision

Measurements with a gas standard 104 Indication of Alphaguard [Bq/m³] 10³ 10² 10¹ ALPHAGUARD (Integration 10 min) 10⁰ 400 200 600 800 1000 Time from reference date [d]



| Verantwortung für Mensch und Umwelt |

(estimated from regression)

 $\lambda = 0,007502 \pm 0,000025 \text{ h}^{-1}$

Slope

Transfer Comparison Device

Procedure for checking linearity and precision



U includes a coverage factor of 2



Transfer Comparison Device

Regular checks of background and instrument settings





Protocols, Information and Documentation

Information about

- Instrument used as transfer comparison device
- Course of the comparison
- Shipment and transport
- Concentration levels

Provision of forms

- for recording calibration procedures and
- for reporting results

Basis for compiling the data and calculating the results



Protocol, Information and Documentation

Radon activity concentrations

	Nominal value	Accepted deviation
1	400 Bq∙m-³	350 Bq·m⁻³ – 450 Bq·m⁻³
2	1000 Bq∙m ⁻³	900 Bq·m ⁻³ – 1100 Bq·m ⁻³
3	6000 Bq∙m ⁻³	5500 Bq·m⁻³ – 6500 Bq·m⁻³



Execution





Participants

	NMI	Montenegro	Goran Vukoslavovic	
	BEV-PTP	Austria	Franz Josef Maringer	
	IRSN	France	Sylvain Bondiguel	
	STUK	Finland	Tuukka Turtiainen	
	SUJCHBO	Czech Rep.	Josef Vošahlik	
	BFKH	Hungary	Norbert Szabó	
	CLOR	Poland	Katarzyna Wołoszczuk	
15 Laboratories -	SSI	Sweden	Jens Jensen	
(12 EU countries +	UNICAN	Spain	Carlos F. Sainz	
Montenegro)	SMU	Slovakia	Matej Krivošík	
	UBB	Romania	Kinga Szacsvai	
	IFIN-HH	Romania		
	UPC	Spain		
	BfS	Germany		
	ENEA	Italy		



Data Assessment

Quantity of Comparison, R_i

Ratio of radon activity concentrations determined by participant and by transfer device:

$$R_i = \frac{\bar{C}_{Participant,i}}{\bar{C}_{transfer,i}}$$

Relative standard deviation of R_i :

$$u_{rel,i}^{2} = \left(\frac{\Delta R_{i}}{R_{i}}\right)^{2} = \left(\frac{\Delta \bar{C}_{Participant,i}}{\bar{C}_{participant,i}}\right)^{2} + \left(\frac{\Delta \bar{C}_{transfer,i}}{\bar{C}_{transfer,i}}\right)^{2}$$

Only statistical uncertainties, no calibration uncertainties



Data Assessment

Uncertainty-weighted mean \overline{R} of all participants

$$\bar{R} = \frac{\frac{R_1}{u_1^2} + \dots + \frac{R_n}{u_n^2}}{\frac{1}{u_1^2} + \dots + \frac{1}{u_n^2}} = \sum_{i=1}^n w_i R_i$$
Normalized weights: $w_i = \frac{1/u_i^2}{\sum_{i=1}^n 1/u_i^2}$
Variance associated with \bar{R} : $\frac{1}{u_i^2(\bar{R})} = \frac{1}{u_i^2} + \dots + \frac{1}{u_n^2}$

 $u^2(R)$



 u_1^2

| Verantwortung für Mensch und Umwelt |

Bundesamt für Strahlenschutz

 u_n^2

Provisional results



Radon Activity Concentration as indicated by the comparison device [Bq.m⁻³]

All uncertainties are given with the extension k=1.



Problems with the common variance (and accordingly standard derivation)

Variance associated with \overline{R} :

$$\frac{1}{u^2(\bar{R})} = \frac{1}{u_1^2} + \dots + \frac{1}{u_n^2}$$

Literature:

... the reciprocal square-root ... becomes too small as the number of participants increases and many labs fall outside the uncertainty interval. [Rukhin, Metrologica 46 (2009)]

The widely used traditional variance estimator ... underestimates the variance ... and the intervals formed by this estimator have poor coverage probability [Zhang, Metrologica 43 (2006)]

The variance estimator seems to be not suitable for quantifying the closeness of agreement (precision) between the participants.



Consistency check of results

Hypothesis H_0 : Results belong to the same basic population Hypothesis H_1 : Results do not belong to the same basic population

Test statistic

$$\chi^{2} = \sum_{i=1}^{n} \frac{(R_{i} - \bar{R}_{i})^{2}}{u_{i}^{2}}$$

n: Number of participants

Hypothesis H₀ has to be rejected, if
$$\chi^2 \ge \chi^2_{n-1;1-\alpha}$$

 $\chi^2_{n-1;1-\alpha}$: Quantile of the χ^2 distribution for the significance level $1-\alpha$



Consistency check of results, Error probability 5% (α =0,05)

Radon Level	п	χ^2		$\chi^2_{n-1;1-lpha}$
400 Bq/m³	9	9,94	<	15,5
1000 Bq/m³	10	5,12	<	16,9
6000 Bq/m³	11	9,44	<	18,3
All	30	24,40	<	42,6

Hypothesis H₀ cannot be rejected: The results are samples of the same population.



The results are mutually consistent. Participants share a common mean value. Deviations from the mean value are normally distributed.



Preliminary Summary

- The comparison is carried out according to relevant guidelines: EURAMET Guide on Comparison; BIPM/CCQM Guidance note on the estimation of a consensus reference value.
- The results are mutually consistent. Not outlier was observed.
- The majority of the calibration facilities shows a closeness of agreement within an interval of ±5% around the common mean values.
- A statistically meaningful quantification of the closeness of agreement is still in progress.





